



Certificate of Traceability

Level I Laboratory Certification

This certifies that Mitsubishi Chemical Medicine Corporation, using Bio-Rad D-10 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#7

The system evaluated was:

Instrument: D-10	Calibrator Lot: AA92280, AA92281	Column Lot: A80100R
Reagent Lot: AA92156, AA92050, AA92421, AA92422, AA92423	Calibrator Assigned Values: 5.4%, 10.6%	

Date of Certification: May 1, 2010

Certification Expires: May 1, 2011

Daniel S. Subis

NGSP Steering Committee Chair

Randie R. Stille PhD.

NGSP Network Coordinator

Sharon Amick

SRL director/ supervisor