



Certificate of Traceability

Level I Laboratory Certification

This certifies that **ICON Laboratory Services China**, using **Trinity Premier Hb9210** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#11**

The system evaluated was:

Instrument: Premier Hb9210	Calibrator Lot: 11300	Column Lot: 10131
Reagent Lot: 10737, 10721, 10724, 10716	Calibrator Assigned Values: 5.5%, 11.9%	

Date of Certification: May 1, 2021

Certification Expires: May 1, 2022

A handwritten signature in black ink, appearing to read 'David Sacks', is written over a horizontal line.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Randie R Little PhD', is written over a horizontal line.

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read 'Rhonda Howard', is written over a horizontal line.

SRL director/ supervisor